

## EU Declaration of Conformity

<b>Product</b>	Laboratory Equipment	<b>File Number</b>	P225
<b>Manufacturer</b>	Cole-Parmer Ltd Beacon Road Stone, Staffordshire ST15 0SA United Kingdom		

This declaration of conformity is issued under the sole responsibility of the manufacturer

**Object of Declaration**

Flame Photometer Range  
(reference the attached list of catalogue numbers)


The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation:

<b>Low Voltage Directive</b>	2006/95/EC
<b>EMC Directive</b>	2004/108/EC
<b>RoHS Directive</b>	2011/65/EC

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

IEC/EN 61010-1:2001	Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements.
IEC/EN 61010-2-010:2003	Particular requirements for laboratory equipment for the heating of materials.
IEC/EN 61326-1:2006	Electrical equipment for measurement, control and laboratory use. EMC requirements. Part 1: General requirements (Class A).

Signed for and on behalf of the above manufacturer

<b>Additional Information</b>	Year of CE Marking: 1996
<b>Place of Issue</b>	Stone, Staffordshire, UK
<b>Date of Issue</b>	October 2009
<b>Authorised Representative</b>	Carl Warren
<b>Title</b>	Technical Manager
<b>Signature</b>	

Range	Catalogue Number	Part Code	Description
FLAME	83055-02	500731	PFP7, Industrial Flame Photometer, 110 V
FLAME	83055-05	500701	PFP7, Industrial Flame Photometer, 230 V
FLAME	NA	500831	PFP7/C, Flame Photometer, 110 V
FLAME	NA	500801	PFP7/C, Flame Photometer, 230 V